

EARL PET/CT

Accreditation 2021

THIS IS TO CERTIFY THAT THE


Klinikum rechts der Isar / TUM Nuklearmedizin Scanner: Siemens Biograph mCT-X TrueV PET PETCT11045

is granted accreditation according to the requirements and specifications indicated in the EARL PET/CT accreditation manual for the following oncology isotope imaging and accreditation program:

^{18}F standards 1, ^{18}F standards 2, ^{89}Zr , ^{68}Ga

PET/CT scanner accreditation ensures similar performance of PET/CT systems within a multicentre setting by harmonising image quality for quantification of PET/CT studies. Accredited PET/CT centers of excellence can compare, exchange and combine PET/CT findings, including SUVs, since data are collected and processed in a standardised manner. The regular submission and approval of QC data, as required by EARL, is a condition for the maintenance of the accreditation.

Date of Print: Q3-2021



Jolanta Kunikowska
President 2021-2022



Ivalina Hristova
Programme Director



EARL PET/CT

Accreditation 2021

THIS IS TO CERTIFY THAT THE

Klinikum rechts der Isar / TUM Nuklearmedizin Scanner: Siemens Biograph Vision 600

is granted accreditation according to the requirements and specifications indicated in the EARL PET/CT accreditation manual for the following oncology isotope imaging and accreditation program:

^{18}F standards 1, ^{18}F standards 2

PET/CT scanner accreditation ensures similar performance of PET/CT systems within a multicentre setting by harmonising image quality for quantification of PET/CT studies. Accredited PET/CT centers of excellence can compare, exchange and combine PET/CT findings, including SUVs, since data are collected and processed in a standardised manner. The regular submission and approval of QC data, as required by EARL, is a condition for the maintenance of the accreditation.

Date of Print: Q3-2021



Jolanta Kunikowska
President 2021-2022



Ivalina Hristova
Programme Director

